List of reference measurement methods/procedures

This file was created on 02 November 2015 from the JCTLM-DB website (http://www.bipm.org/jctlm/)
Your search criteria: Reference measurement methods/procedures; JCTLM DB identification number: Lab 25

IFCC reference measurement procedure (37 °C) for AST	
Analyte	aspartate aminotransferase (AST)
Applicable matrice(s)	lyophilized, fresh, or frozen human serum or plasma
Full description of technique(s)	Kinetic spectrophotometry
Quantity	Catalytic amount
Applicable range	0.028 µkat/l to 4.17 µkat/l
Expected uncertainty (level of confidence 95%)	2.2% to open
Reference(s)	*Clin. Chem. Lab. Med.*, 2002, **40**, 725-733
Comparability assessment study(ies)	In press, data available by request at lothar.siekmann@ukb.uni-bonn.de
JCTLM DB identification number	NRMeth 68