

List of reference measurement methods/procedures

This file was created on 02 November 2015 from the JCTLM-DB website (<http://www.bipm.org/jctlm/>)
Your search criteria: Reference measurement methods/procedures; JCTLM DB identification number: Lab 25

▶ IFCC reference measurement procedure (37 °C) for LDH	
Analyte	lactate dehydrogenase (LDH)
Applicable matrice(s)	lyophilized, fresh, or frozen human serum
Full description of technique(s)	Kinetic spectrophotometry
Quantity	Catalytic amount
Applicable range	0.068 μ kat/l to 10 μ kat/l
Expected uncertainty (level of confidence 95%)	2.2% to open
Reference(s)	<i>Clin. Chem. Lab. Med.</i> , 2002, 40 , 643-648
Comparability assessment study(ies)	<i>Clin. Chem. Lab. Med.</i> , 2002, 40 , 739-745
JCTLM DB identification number	NRMeth 66