



Bilirubin EQA

INTENDED USE

Weqas *Bilirubin EQA* samples are for in-vitro diagnostic use as an external quality assessment material for testing of Total Bilirubin and Conjugated Bilirubin.

SUMMARY

External Quality Assessment (EQA), or proficiency Testing (PT) is an essential part of providing quality laboratory diagnostic services, and participation in EQA is required for laboratory accreditation to ISO 15189 and ISO 17025. EQA is the inter laboratory comparison and performance evaluation that extends throughout all phases of the healthcare diagnostic testing cycle.

PRODUCT DESCRIPTION

1.0 mL volume is supplied in sterile plastic tubes. The material has been prepared from sterile human serum, with an antibiotic added to maintain sterility. The serum is dispensed then stored at -20°C until dispatch.

STORAGE AND STABILITY

Unopened samples are stable for two (2) weeks in a refrigerator at 2-8°C.

Samples should be protected from light.

The samples should be assayed on the day you receive them. If you are unable to assay the samples immediately upon receipt, please store at 2-8°C until analysis. Samples stored at 2-8°C prior to analysis must be brought to ambient temperature (18-30°C) and assayed within 60 minutes.

Samples should be stored at -20°C for long term storage.

STORAGE AND PROCEDURE FOR SAMPLES RECEIVED BY BULK DISPATCH

The samples should be stored at -20°C on receipt. At the start of the return window for each distribution thaw the relevant samples at ambient temperature (18-30°C) for 24 hours (maximum 48 hours) and assay as described in PROCEDURE.

Please ensure that the samples are well mixed before analysis.

PROCEDURE

The samples should be treated the same as patient specimens and run in accordance with the instructions accompanying the test system being used.

1. Mix the sample well by gently inverting 5 to 6 times.
2. Wear gloves and handle the sample as a normal patient sample.
3. Safely dispose of excess sample in accordance with local waste policy guidelines.

Always wear gloves to avoid contamination.

LIMITATIONS OF PROCEDURE

The *Bilirubin EQA* samples should be analysed according to the instructions within this document. If there is evidence of microbial contamination in the product, discard the vial.

The *Bilirubin EQA* samples require storage as described in STORAGE AND STABILITY and handling as described in PROCEDURE.

Accurate and reproducible results are dependent upon properly functioning instruments and reagents and the use of correct procedures.

ANALYTES COVERED

The Weqas *Bilirubin EQA* samples cover a relevant pathological and analytical range for the analytes listed below.

Analyte
Total Bilirubin
Conjugated (Direct) Bilirubin

! CAUTION !

Human source material. Treat as potentially infectious.

Each unit of whole blood used in the manufacture of this product has been tested and found to be non-reactive for the presence of HBsAg and antibody to HIV-1/2, HCV and HIV-1 Ag.

Although every effort is made to ensure that the material is free from any known infectious agent, the samples should be handled as for clinical specimens.