

# Weqas

GLOBAL PROVIDER OF QUALITY  
IN DIAGNOSTIC MEDICINE



## INTERPRETATION OF LABORATORY EQA REPORTS





Published: 06/22

# Contents

<b>1 Statistical Analysis - Quantitative Programmes</b> .....	<b>4</b>
1.1 Target Value Assignment and Traceability .....	5
1.1.1 Reference Values .....	5
1.1.2 Uncertainty .....	5
1.2 Comparability Factors .....	5
1.3 Scoring System .....	6
1.4 Performance Criteria .....	7
1.5 Minimum Analytical Performance Standards (MAPS) .....	8
1.6 The Weqas Report .....	10
1.6.1 Measurements of Imprecision .....	11
1.6.2 Measurements of Inaccuracy .....	12
1.6.3 Bias Plots .....	13
1.7 Problem Solving .....	16
1.8 Other Reports .....	23
<b>2 Statistical Analysis - Qualitative Programmes</b> .....	<b>28</b>
2.1 Target Value Assignment .....	28
2.2 Scoring System .....	28
2.3 The Weqas Report .....	29

# 1. Statistical Analysis - Quantitative Programmes

THE FOLLOWING APPLIES TO ALL LABORATORY PROGRAMMES AND THE POCT PROGRAMMES LISTED IN THE FOLLOWING TABLES. POCT PROGRAMMES NOT LISTED ARE COVERED IN A SEPARATE DOCUMENT. QUALITATIVE PROGRAMMES ARE COVERED IN SECTION 2.

The linear panel of samples used in most Weqas Programmes allow the evaluation of *inaccuracy, within run imprecision and between batch imprecision.*

**Table 1 - Laboratory Programmes**

Weqas Programme Title	Additional Sub Programmes / Comments
Serum Chemistry	Serum Indices interference studies
Bilirubin	
Lipid	
ED Toxicology	Whole Blood Ethanol. *Ethylene Glycol and Methanol.
Serum Indices	
Blood Gas and Co-oximetry	Haematocrit. Oxygen saturation and calculated parameters. b123 Co-oximetry.
HbA1c	
Ammonia	
Endocrine	Macroprolactin
Haematinics	Iron overload
Cardiac Marker	Semi-Quantitative
BNP and NT pro BNP	
Homocysteine	
Bile Acids	
Urine Chemistry	Acidified samples for Ca, Mg and Phosphate
Urine Oxalate and Citrate	
Serum hCG	Qualitative and Quantitative Serum hCG
Porphyrin	Includes Quantitative and Qualitative Urine, plasma, faeces and clinical cases
Serum ACE	
CRP	Includes hsCRP
TDM	
Whole Blood Immunosuppressants	
Drugs of Abuse	
Quantitative Faecal Hb	
pH Meter	
Procalcitonin	

**Table 2 - POCT Programmes**

Weqas Programme Title	Additional Sub Programmes / Comments
Pregnancy Testing	Qualitative Urine and Serum Programmes
Blood Gas / Co-oximetry	Offered with simplified reports in Lab Programme
Bilirubin	Offered with simplified reports in Lab Programme for Bilirubinometer / Blood Gas analysers
POCT HbA1c	Bimonthly Programme offered with simplified reports in Lab Programme
POCT Cardiac Marker	Plasma CM available for Triage meters Serum CM available for other POCT devices as part of Lab Scheme
POCT HIV	
Pre Term Labour Markers	Foetal fibronectin. Phosphorylated IGFBP-1. IGFBP-1.
POCT BNP	Plasma BNP available for Triage meters
POCT Creatinine	
Drugs of Abuse	Offered with simplified reports in Lab Programme
POCT CRP	
POCT Hb	
POCT INR	

<sup>+</sup> Pilot (Not Accredited)

## 1.1 Target value assignment and Traceability

Statistical methods that are robust to outliers complying with ISO 13528:2015: Statistical methods for use in proficiency testing by interlaboratory comparison are used. For each analyte for each sample the overall Robust mean and standard deviation is calculated using Algorithm A with iterated scale.

Methods are grouped into broad method groups based on the principle of the method, e.g. Glucose Method 1 = Glucose Oxidase, Method 2 = Hexokinase as well as the platform (analyser) type. The Robust method mean and analyser mean are calculated using Algorithm A as above. Each laboratory's results are compared against target values using linear regression analysis to give a measure of systematic error. The target value can either be:

- 
- Reference value – where the sample is measured using a validated reference method traceable to a high metrological order or by gravimetric measurement.
  - Method mean – used if no reference target values are available and the number of participants using the method  $\geq 8$
  - Overall mean – used if no reference target values are available and the number of participants using the method  $< 8$

Analyser mean – this is provided on the report for information only and is not used to calculate the target value unless the analyser is regarded as sufficiently different to other systems to justify its own method group e.g. Ortho Vitros.

### 1.1.1 Reference Values

The **HDL Cholesterol** target values are assigned using the CDC Abel-Kendall reference method in an approved CDC network laboratory.

For **HbA1c** the target values are assigned using the IFCC methods in an approved IFCC (NGSP) reference laboratory.

**All other Reference values are assigned by the Weqas Reference Laboratory.**

### 1.1.2 Uncertainty

The combined standard uncertainty of the reference target value is calculated from the *ISO Guide to the Expression of Uncertainty in Measurement*.

$$\text{Combined Standard Uncertainty} = \sqrt{\{(\text{U}_{\text{sample}})^2 + (\text{U}_{\text{std}})^2 + (\text{U}_{\text{SRM}})^2\}}$$

Where

- U<sub>sample</sub> = uncertainty associated with sample precision
- U<sub>std</sub> = uncertainty associated with standard preparation
- U<sub>SRM</sub> = uncertainty associated with the SRM


An estimate of the uncertainty of the Robust mean is calculated from:

$$\text{Estimated Uncertainty} = \frac{1.25 \times \text{SD}}{\sqrt{n}}$$

## 1.2 Comparability Factors

This is used for multimodal data where a wide variation is observed for the overall consensus mean due to the widely different methods used. Typically, this would be used for enzymes, (as activity rather than concentration is measured), Ammonia (wet and dry chemistry systems) and Troponin I (no standardisation). A method specific comparability factor (CF) is calculated for each method by analysing the method data using linear regression analysis against a peer reference method (i.e. IFCC for enzymes, GLDH method for Ammonia and Beckman AccuTnl for Troponin I.). An example for Troponin is given in Figure 1. The results for each laboratory are then adjusted using the CF. Each laboratory's results can therefore be compared with their own method group, the peer reference method and directly compared with the overall mean of all groups. The CF's for each scheme are available on request. Where applicable, the recommended IFCC methods have also been set up in the Weqas Reference Laboratory to give definitive values.

**Figure 1 - Relationship between Troponin I methods and the calculation of method specific CF**



### 1.3 Scoring System

For each analyte at each sample point the standard deviation index (SDI) is calculated. This is calculated as: (laboratory result – target value) / Weqas SD. In some EQA Programmes this is known as the Z score. The target value is described in section 1.6.

For each analyte the average SDI is calculated to give an **analyte SDI**. This is calculated as the sum of the absolute numerical values of the individual SDI scores divided by the number of scores. The positive and negative signs are not included in the calculation as this will mask poor performance. An acceptable average analyte SDI does not guarantee acceptable performance across the analytical range and the individual scores must be looked at.

**Table 5 - Interpretation of Scoring System Based on SD Index**

less than 1	Good - all points within $\pm 1$ SD	
1 - 2	Acceptable	
greater than 2	Unacceptable - Laboratory needs to evaluate the analyte	

The SDI is an index of Total error and will include components of both inaccuracy and imprecision.

#### *Running Score of Lab SDI*

This gives a general overview of performance over time. The median (50th centile), and worst SDI scores (97.5th centile) for all laboratories are given for comparison.


## 1.4 Performance Criteria

Laboratories must ensure that the analytical quality attained is appropriate for the needs of the clinical service. It is therefore essential that EQA performance criteria should also reflect clinical need. A hierarchical strategy to establish analytical goals was proposed at the European Federation of Laboratory Medicine in Milan in 2014 and is summarized below.

- **Model 1. Based on the effect of analytical performance on clinical outcomes.** This model is the most rationale since it is based on the actual clinical outcome; however, in practice it is applicable only to a few tests since it is difficult to show the direct effect of laboratory tests on medical outcome.
- **Model 2. Based on components of biological variation of the measurand.** This model seeks to minimize the ratio of the analytical noise to the biological signal. Its applicability can however be limited by the validity and robustness of the data on biological variation.
- **Model 3. Based on the state of the art.** This model is the one where data is most easily available. It is linked to the highest level of analytical quality achievable with the currently available techniques.

The models higher in the hierarchy are to be preferred to those at the lower level. Different strategies have been applied to the different analytes in each scheme based on what is achievable. If the biological goals are not achievable, the analytical performance criteria are based on current “state of the art” of the methods. These “state of the art” precision profiles are calculated over several batches over a wide pathological range. The relationship between SD (or CV%) and the analytical concentration is calculated from the line of best fit (often polynomial). Figure 2 shows an example for Serum Creatinine. These analytes are reviewed every 2 years and approved by the Steering Committee.

Figure 2 - Precision profile for Serum Creatinine



## 1.5 Minimum Analytical Performance Standards (MAPS)

MAPS is a National Quality Assurance Advisory Panel (NQAAP) initiative endorsed by the Professional bodies; the Royal College of Pathologists, ACB, ACP and IBMS. Five analytes have been included in the first pilot: Cholesterol, HDL, Glucose, Creatinine and HbA1c.

MAPS is based on the European Biological Variability Data now hosted by the European Federation of Clinical Chemistry and Laboratory Medicine (EFLM) at <https://biologicalvariation.eu/>

For each test the standard has been defined against a critical diagnostic value and reference method at which the MAPS should be assessed, and provide values for Bias, Imprecision (CV) and Total Error (TE).

**Table 6 - MAPS Phase 1 Analytes**

Analyte	Criteria level	TE %	Bias %	CV %	Sigma	Ref Method
Cholesterol	5 mmol/l	8.5	4	2.7	1.67	CDC
HDL	1 mmol/l	15.9	10	3.6	1.64	CDC
Glucose	7 mmol/l	6.9	2.2	2.9	1.62	ID-GCMS
Glucose	2 mmol/l	10				ID-GCMS
HbA1c	50 mmol/mol	7.7	3.6	2.5	1.64	IFCC
Creatinine	75 umol/l	9.5	5	2.7	1.67	ID-GCMS

For these MAPS, TE = (1.65\*imprecision)+inaccuracy

### How does Weqas calculate these parameters and provide a MAPS score?

Weqas has combined the MAPS specifications into a single score called the Sigma Score.


$$\text{Sigma} = [(TE_{\text{maps}} - \text{bias}_{\text{obs}}) / s_{\text{obs}}]$$

Where  $TE_{\text{map}}$  is the Total allowable error as defined by MAPS

$Bias_{\text{obs}}$  is the laboratory Bias at the critical level and is calculated from the linear regression analysis,  $y = mx + c$  (uses 4 results, can be from previous dist). This can only be calculated if the true bias to reference method is known.

$s_{\text{obs}}$  is the within run CV(%) and is calculated from the  $Sy.x$ . (see section 1.6.1)

**Figure 3 - Example of Sigma Score Calculation - Distribution L299 HDL Cholesterol**



## From the above example for HDL,

### **Bias**

The bias is calculated from the linear regression analysis of the laboratory results against the CDC Reference Laboratory's target values.

From the linear regression analysis equation,  $y = mx + c$ , the bias is calculated at the critical level ( $x$ ), which for HDL is 1 mmol/L.

When  $x=1$  then  $y=0.99(*1)+0.119 = 1.109$ , (intercept rounded up on report to 0.12).

Therefore, bias =  $(y-x)/x * 100 = (1.109-1)/1 * 100 = 10.9\%$

Interpretation - This is higher than the MAPS allowable bias of 10% and needs action.

### **Imprecision**

Laboratory within run Imprecision,  $S_{yx} = 0.032$  mmol/L

$CV = (S_{yx}/x) * 100 = 0.032/1 * 100 = 3.2\%$

Interpretation – This is within the MAPS allowable CV of 3.6% and is therefore acceptable.

### **Sigma Score**

$\text{Sigma} = [(TE_{\text{maps}} - \text{bias}_{\text{obs}}) / S_{\text{obs}}]$

For HDL  $TE_{\text{maps}} = 15.9\%$

Therefore  $\text{Sigma} = (15.9 - 10.9) / 3.2 = 1.56$

The MAPS allowable Sigma is calculated from:

$\text{Sigma}_{\text{min}} = (TE_{\text{maps}} - \text{Bias}_{\text{maps}}) / S_{\text{maps}}$

$\text{Sigma}_{\text{min}} = (15.9-10)/3.6 = 1.64$

During the pilot, the only additional information displayed on your report will be the Sigma score.

Laboratory performance that does not meet the MAPS criteria will be highlighted in red.

### **How does MAPS Score affect poor performance surveillance?**

During the pilot, the SDI score will remain as the index for poor performance surveillance and the existing analytical specifications for the SDI calculation will remain unchanged. The Sigma score will however be used to identify methods that do not comply with MAPS and the manufacturers contacted.

## 1.6 The Weqas Standard Report - An annotated version is provided in Figure 6a and 6b.

The Report outlines the Laboratory Code, Section Code, Distribution Code and sample numbers. The current method code is printed against each analyte. The following table outlines the parameters covered in the Weqas report.

Reported Results	Results as submitted on the "Result Entry form"
Method corrected results	Results adjusted if a method CF is used. Lab result / method CF
Method mean	Estimation of the method mean using a robust algorithm
Method SD	Estimation of the Method SD using a robust algorithm
Analyser mean	Estimation of your analyser group mean using a robust algorithm
Analyser SD	Estimation of your analyser group SD using a robust algorithm
Number of results	Number of results in your method group
Overall mean	Estimation of the overall mean using a robust algorithm
Weqas SD	SD used to calculate SDI and given in graphical representation - fixed for a given level of analyte. Performance criteria = target value $\pm$ 2*Weqas SD
Overall number	Number of reported results
SDI	(Laboratory result – target value)/ Weqas SD
Reference values	Target values using validated reference methods
Uncertainty of target value	The standard uncertainty of the target value is calculated from the <i>ISO Guide to the Expression of Uncertainty in Measurement</i> .
Non scoring reference value	For information only, used when the reference method procedure gives very different results to routine methods.
Sigma score	Your score based on MAPS criteria.
Analyte SDI	Your Average SDI for the analyte
Overall Section SDI / Lab SDI	Overall SDI for your section or Lab.
Previous SDI	Accumulator of previous SDI scores for your lab
Median All Laboratory SDI	Median (50th centile) SDI for all laboratories for this distribution. SDI < Median indicates good score (top 50% of labs).
97.5th Centile	SDI poor performer indicator. SDI > than this value indicates poor lab score (worst 2.5% of labs).
Correlation coefficient	This is used as an index of within run imprecision, the wider the deviation from 1.000, the wider the scatter of results about the line of best fit.
Standard deviation of the residuals	This is used as an index of within run imprecision, and is provided in the units of the analyte. It gives an indication of standard deviation across the range of samples.
Imprecision score	This is derived from the correlation coefficient.
Linear regression equation	This is used as an index of inaccuracy. The slope should be as close to 1.0 and the intercept should be as close to 0. It provides a measurement of agreement between your results and the target value over a range of samples.

### 1.6.1 Measurements of Imprecision

The Coefficient of Linear Correlation and the Standard Deviation of the Residuals gives a measure of the dispersion of the points about the best fit line and is therefore an index of precision. The Imprecision Score is derived from the correlation coefficient.


- **Standard Deviation of the Residuals ( $S_{y.x}$ )**

The equation for the  $S_{y.x}$  is:

$$\frac{\sqrt{\sum dy.x^2}}{d.f}$$

where  $y$  = observed value,  $d.f$  = degrees of freedom and  $\tilde{y}$  is the value on the line of best fit

**Figure 4 - Example of Calculation of the  $S_{y.x}$**



Target	Lab result	Line of fit	Deviation	Dev <sup>2</sup>
$x$	$y$	$\tilde{y}$	$dy.x = y - \tilde{y}$	$dy.x^2$
111	108	109.9	-1.9	3.62
123.5	128	123.3	4.7	21.97
135.7	136	136.4	-0.4	0.16
148	144	149.6	-5.6	31.30
160.3	166	162.8	3.2	10.31
	<i>slope</i>	<i>1.07</i>	$\sum dy.x$	$\sum dy.x^2$
	<i>int</i>	<i>-9.17</i>	0.0	67.36
	<i>r</i>	<i>0.9812</i>		
	<i>IS</i>	<i>187.7</i>		
	$S_{y.x}$	<i>4.74</i>		

### The Coefficient of Linear Correlation (r):

The equation for the correlation coefficient is:

$$\text{Correl}(X, Y) = \frac{\sum(x-\bar{x})(y-\bar{y})}{\sqrt{\sum(x-\bar{x})^2 \sum(y-\bar{y})^2}}$$

- Imprecision score (IS)

The equation for the IS:

$$\text{IS} = (1 - r) * 10,000$$

**Table 7 - Interpretation of “r” value and imprecision score**

'r' value	Imprecision score	Interpretation
0.9990 to 1.0000	0 to 10	Good
0.9850 to 0.9989	11 to 150	Acceptable to Warning level
< 0.9850	> 150	Unacceptable (including Curvilinear Data)

### 1.6.2 Measurements of Inaccuracy

The Linear Regression Analysis of the laboratory results ( $y$ ) against the target value ( $x$ ) is used as an index of inaccuracy. **Linear regression** produces the slope of a line that best fits a single set of data. The equation  $y = mx + c$  algebraically describes a straight line for a set of data with one independent variable where  $x$  is the independent variable,  $y$  is the dependent variable,  $m$  represents the slope of the line, and  $c$  represents the y-intercept.

- The accuracy of the line calculated depends on the degree of scatter in your data. The more linear the data, the more accurate the model. Weqas uses the method of least squares for determining the best fit for the data. The calculations for  $m$  and  $c$  are based on the following formulas:

$$m = \frac{\sum(x-\bar{x})(y-\bar{y})}{\sum(x-\bar{x})^2} \qquad c = \bar{y} - m\bar{x}$$

A deviation from a slope ( $m$ ) of 1.00 indicates possible systematic proportional error.

The intercept ( $c$ ) gives an indication of the systematic absolute (blank) error.

$y = mx + c$  is not calculated where the 'r' value is below 0.9.

### 1.6.3 Bias Plots

The bias plot gives a graphical representation of each laboratory's values compared with the "target" values. The "x" axis line represents the "target" values. The "y" axis has a scale that spans  $\pm 3$  SD from this line, and the 2 SD limits are marked on the graph (...) Standard deviation limits used in the report are calculated from the analyte performance criteria.

#### Left hand graph

This represents the current distribution. "x" indicates the laboratory bias at each level of analyte; "o" indicates the method mean bias and "□" the instrument mean bias. The bar lines relate to the  $\pm 2$  SD limits around the method mean. At the right hand side of each graph the relationship between the laboratory's results and the target value is expressed as a straight line equation, 'y = mx + c'. The Coefficient of Linear Correlation, (r) the Standard Deviation of the Residuals (Sy.x), and the Imprecision Score (IS), are also given.

#### Right hand graph

This provides a cumulative bias plot of the data over 6 distributions and shows a graphical display of the between batch imprecision.

Figures 5a and 5b - Bias Plots




Figure 6a

**Standard Report Format - Easy to read cumulative reports**




QA Officer  
Biochemistry  
St Elsewhere  
Inacity  
Someshire  
KK1 4ZX

**Manager's performance summary sheet.**

**Total Error** - measured as **Standard Deviation Index**. -includes components of inaccuracy and imprecision.

This Distribution	
Overall Lab SDI:	0.36
Median All Laboratory:	0.52
97.5th centile:	1.21



All SDI Ranges	
< 1	Good
1 - 2	Acceptable
> 2	Poor

---- Median    ● Lab SDI    — 97.5th

**Section SDI scores for this distribution**

Section	Harris (700)	Tweed (800)
Overall	0.41	0.31
Cholesterol	0.46	0.30
Triglyceride	0.42	0.35
HDL Cholesterol	0.22	0.20
LDL Cholesterol	0.55	0.39

For each sample for each analyte:  
**SDI = (Lab result – \*Target value) / Weqas SD\*CF**

**Running Score of Lab SDI** - gives a general overview of performance over time. The best and worst SDI scores for all laboratories are given for comparison.

**Analytical goals (Weqas SD)** are based on either:  
•Precision profiles - reflecting the "state of the art".  
•Clinical decision goals -e.g. cholesterol  
•Biological variation- e.g. HbA1c

Analyte SDI for each Section at a glance. Colour coded for performance.

SDI Code	Meaning
N/A	Not enrolled for this analyte
?	Analyte enrolled but no results returned
N/S	This analyte not scored
**	SDI score greater than 2

**Hierarchy of target values**

- 1 Reference method (if available for all samples)
- 2 Method mean (if n > 8)
- 3 Overall mean

Comments:  
Information relating to this distribution can be displayed in this text box

Figure 6b

Analyte Specific data sheet - detailed assessment of performance for each analyte

**Reference values.**  
If Reference values are available they are used to calculate the SDI and replace the overall mean as the target value on the bias plot.

Scheme: Lipid. Distribution Code: L364. Distribution Date: 26/03/18. Final Report issued: 20/04/18					
Cholesterol (mmol/l)	1	2	3	4	Analyte SDI
Reported Result	6.80	2.50	4.60	4.80	
Method Corrected Result	6.800	2.500	4.600	4.800	
Cholesterol oxidase	Mean	7.092	2.617	4.876	4.963
	SD	0.225	0.093	0.183	0.169
	Number	174	174	174	174
	Uncert.	0.0214	0.0088	0.0174	0.0161
Cobas C Module	Mean	7.006	2.587	4.811	4.901
	SD	0.123	0.075	0.112	0.106
	Number	98	98	98	98
	Uncert.	0.0155	0.0094	0.0142	0.0134
Overall	Mean	7.081	2.615	4.868	4.953
	SD	0.233	0.090	0.189	0.174
	Number	179	181	179	181
	Uncert.	0.0218	0.0084	0.0176	0.0161
Reference Values CDC		7.038	2.606	4.867	4.963
Ref. Value Uncertainty		0.0140	0.0000	0.0000	0.0000
Non-scoring Reference Values ID-GCMS		7.110	2.590	4.850	4.970
WeQas SD		0.301	0.111	0.207	0.210
<b>SDI</b>		<b>-0.79</b>	<b>-0.95</b>	<b>-1.29</b>	<b>-0.77</b>
<b>Sigma Metrics</b>					
Critical Level 1: 5.0 mmol/l					
Minimum Acceptable score	1.67	Critical Level 1 Sigma score			3.8
MAPS Allowable TE	8.5%				
MAPS Allowable bias %	4.0%	Lab  bias  %			3.9%
MAPS Allowable CV %	2.7%	Lab CV %			1.2%


Please note: Linear regression uses CF corrected data.

**Total Error**

SDI is a measurement of your total error and will include both inaccuracy and imprecision


**This Distribution L364**  
Your average analyte SDI for the 4 samples is 0.95

**Previous SDI**




**Sigma Score using MAPS criteria.**

**This Distribution L364**



**Previous Distributions**



**Analytical goals**  
2 x Weqas SD

**Bias Plot (Bland-Altman)**

**Current distribution** (from data in table) including lab, method and instrument performance.

Right hand plot shows **Previous distribution** results, including current distribution

Precision and accuracy indicators

**Precision**

This Distribution L364	Previous Distributions	L363	L362	L361	L360	L359	L358
Sy.x = 0.061 mmol/l	Sy.x	0.046	0.092	0.098	0.058	0.074	0.034
IS = 4	IS	6	9	11	6	11	3

Sy.x is the average deviation from the best fit line and is an index of scatter.

**Precision Key**

IS score	Interpretation
0 to 10	Good
11 to 150	Acceptable to Warning level
> 150	Unacceptable (including Curvilinear Data)

**Accuracy**

This Distribution L364	Previous Distributions	L363	L362	L361	L360	L359	L358
Systematic proportional error (calibration) -2.96%	Proportional (%)	-4.76	-0.40	2.58	0.31	3.60	-0.28
Systematic constant error (blank) - 0.049 mmol/l	Constant (mmol/l)	0.121	0.020	-0.153	-0.005	-0.175	-0.061

Bias includes components of proportional and constant errors. A proportional bias suggests an error of calibration whilst a constant bias suggests a blank error. Mixed errors will include significant components of both.

## 1.7 Problem Solving

The following pages include a rule based problem solving guide. The power of prediction identifies the cause of the problem often before the analysis is out of control, i.e. outside  $\pm 2$  SD. The sensitivity of statistical parameters to different types of error is explained. The types of errors are shown graphically in Figure 7. Problem solving flow charts (Figures 8a and 8b) also allow for a simplified procedure for identifying problems and verifying corrective action.

**Table 8 - Sensitivity of Statistical Parameters to Different Types of Errors**

	<i>Type of Error</i>				
	Imprecision	Inaccuracy (Systematic)			
	<i>Random</i>	<i>Curvilinear</i>	<i>Proportional</i>	<i>Mixed</i>	<i>Constant</i>
Slope, m	No	Yes	Yes	Yes	No
y intercept, c	No	Yes / No	No	Yes	Yes
Standard error, $S_{y.x}$	Yes	Yes	No	No	No
Corr. coefficient, r	Yes	Yes	No	No	No

**Imprecision:** Errors of imprecision should be corrected first. A small random error is acceptable.

**Inaccuracy:** Systematic errors can be eliminated by appropriate improvement in methodology. A small systematic error is tolerable. This depends on the clinical usefulness of the method.

**If your results show an error: look at the Problem Solving Guide flow diagram and identify the error.**

On the Bias Plot the  $y = mx + c$  assumes a linear relationship between the laboratory results and the Ideal Line. For this reason large random errors, identified as an  $IS > 150$  or a wide  $S_{y.x}$  will invalidate this equation. **A line drawn through the points will aid in identifying the type of error.**

### Start by asking the question - Is it Imprecision?

Check for causes of imprecision in the following order:

- Exclude apparent imprecision due to curvilinear data.
- Exclude clerical errors (blunder error).
- Check for causes of imprecision, e.g. inexperienced operators (analysts), faulty equipment, inappropriate methods.

**Once you are happy with your analytical precision you can then look for causes of inaccuracy.**

## Is it Inaccuracy?

Inaccuracy can be due to:

- *Curvilinear data*: Reagent or standard deterioration.
- *Systematic constant*: Usually blank due to reagent, serum or instrument zero.
- *Systematic proportional*: Usually due to calibration, standards.
- *Mixed systematic*: On one point calibration with a cross-over at or near a calibration point (pivoting about calibration point), check zero calibration point, i.e. reagent blank, serum blank, instrument zero and then follow guide as for proportional systematic error. For a two point multi calibration with cross-over at or near one point, check other calibrators and/or zero point.

Figure 7 - Errors in Accuracy




Figure 8a - Problem Solving Flow Chart




Figure 8b - Problem Solving Flow Chart






Figure 9a - Bias Plot - With Explanation

**Sodium**  
mmol/l




$y = 0.9x + 9.6$   
 $r = 0.9995$   
 IS = 5  
 $Sy.x = 0.63$

**Potassium**  
mmol/l




$y = \text{slope not calculated}$   
 $r = 0.8826$   
 IS = 1174  
 $Sy.x = 1.10$

**Chloride**  
mmol/l



$y = 1.01x - 4.01$   
 $r = 0.9995$   
 IS = 5  
 $Sy.x = 0.49$


**Bicarbonate**  
mmol/l



$y = \text{slope not calculated}$   
 $r = 0.8484$   
 IS = 1516  
 $Sy.x = 2.41$


Figure 9b - Bias Plot - With Explanation

**Urea**  
mmol/l




$y = 1.00x - 0.14$   
 $r = 1.000$   
 $IS = 0$   
 $Sy.x = 0.03$

**Glucose**  
mmol/l




$y = 1.09x - 1.06$   
 $r = 0.9991$   
 $IS = 9$   
 $Sy.x = 0.40$

**Calcium**  
mmol/l



$y = \text{slope not calculated}$   
 $r = 0.9810$   
 $IS = 190$   
 $Sy.x = 0.09$

**ALP**  
IU/L



$y = 0.57x + 0.6$   
 $r = 0.9999$   
 $IS = 1$   
 $Sy.x = 0.76$

## Use of the Problem Solving Guide for Identifying Errors in Lab Report

### Sodium (Figure 9a)

- [1] Imprecision - satisfactory [5] Inaccuracy - identify error  
[6]  $m = 0.9$ ,  $c = +9.6$  mmol/L - OK at 100 mmol/L  
2.5% negative bias at 130 mmol/L, 4% negative bias at 160 mmol/L  
Error - mixed. Two points calibration at 110 and 160 mmol/L  
**Cause - incorrect values for 160 mmol/L calibration**

### Potassium (Figure 9a)

- [1] Imprecision - unsatisfactory,  $r = 0.8826$ ,  $Sy.x = 1.1$  mmol/L  
[2] Not curvilinear  
Error - blunder  
**Cause - clerical error, samples 4 and 5 were transposed**

### Chloride (Figure 9a)

- [1] Imprecision - satisfactory [5] Inaccuracy - identify error  
[6]  $c = -4.0$  mmol/L  
Error - systematic absolute. Results low by 4.0 mmol/L over the whole range  
**Cause - incorrect serum blank compensation**

### Bicarbonate (Figure 9a)

- [1] Imprecision - unsatisfactory,  $r = 0.8484$ ,  $Sy.x = 2.4$  mmol/L  
[2] Not curvilinear  
Error - random  
**Cause - faulty syringe on instrument**

### Urea (Figure 9b)

- [1] Imprecision – satisfactory [5] Inaccuracy - satisfactory

### Glucose (Figure 9b)

- [1] Imprecision - satisfactory [5] Inaccuracy - identify error  
[6]  $m = 1.09$ ,  $c = -1.06$  mmol/L  
Error - mixed. One point calibration at 9 mmol/L  
**Cause - incorrect instrument zero**

### Calcium (Figure 9b)

- [1] Imprecision - unsatisfactory,  $r = 0.9810$ ,  $Sy.x = 0.09$  mmol/L  
[2] Error - curvilinear data  
**Cause - time expired reagents**

### ALP (Figure 9b)

- [1] Imprecision - satisfactory [5] Inaccuracy - identify error  
[6]  $m = 0.57$   
Error - systematic proportional error. Results low by 41% over the whole range  
**Cause - incorrect method group classification. The lab was using AMP not DEA buffer.**

Case studies of EQA reports including interpretive comments are available to download from our website. Please use the following link and search for 'Case Studies'.

<http://www.weqas.com/resourcelibrary/>

## 1.8 Other Reports

A number of additional reports are provided which can be accessed online.

### Figure 10 - Analyser reports

Detailed performance reports allowing comparison with all other returns in your own method and instrument group can be accessed via links in the individual analyte report pages.

#### Reported Results for all sections with Instrument Means and SDs

Distribution: OI  
 Distribution Date: 1 Aug, 2011  
 Analyte: Magnesium  
 Method: Magon / Xylidyl blue  
 Instrument: Advia 1200/1650/1800/2400

Distribution Code : OI Sent on: 1/08/11					
Magnesium (mmol/l)	1	2	3	4	
AAE	1.58 *	1.56	0.55	1.13	
BY	1.71	1.57	0.55	1.21	
BY	1.67	1.56	0.50	1.15	
CH	1.79	1.63	0.46	1.18	
CH	1.79	1.64	0.50	1.20	
CH	1.75	1.62	0.52	1.19	
DB	1.72	1.60	0.40	1.14	
DB	1.62	1.52	0.47	1.13	
DS	1.69	1.55	0.50	1.16	
DS	1.68	1.55	0.48	1.14	
EB	1.58 *	1.50 *	0.46	1.09	
EB	1.74	1.62	0.52	1.19	
EC	1.77	1.63	0.51	1.19	
EK	1.77	1.64	0.53	1.21	
EK	1.79	1.64	0.51	1.21	
EQ	1.74	1.59	0.59	1.26	
EQ	1.76	1.60	0.52	1.22	
KH	1.81	1.66	0.51	1.21	
KH	1.73	1.61	0.50	1.18	
KJ	1.72	1.58	0.44	1.18	
KJ	1.72	1.59	0.50	1.19	
KK	1.80	1.66	1.21 *	0.52 *	
KK	1.80	1.64	0.49	1.20	
KK	1.83	1.67	0.52	1.23	
NW	1.77	1.62	0.52	1.21	
NW					
NW	1.78	1.66	0.53	1.20	
SS	1.80	1.66	0.52	1.23	
Overall	Mean	1.702	1.573	0.502	1.169
	SD	0.063	0.055	0.023	0.041
	CV	3.73	3.53	4.60	3.48
	Number	223.00	222.00	232.00	232.00
	Reference Value FAAS / FAES	1.698	1.572	0.484	1.149
	Reference Value NS				
Instrument Specific Data					
Magon / Xylidyl blue	Mean	1.743	1.609	0.509	1.187
	SD	0.052	0.037	0.036	0.036
	CV	2.96	2.29	7.05	3.05
	Number	36.00	36.00	37.00	37.00
Key: Red - Outside Range. * - Instrument Outlier					

## Figure 11 - Method Summary reports

A selection of Method summary reports are e-mailed with your PDF reports and attached to your 'report ready' email notification. Additional summaries for all methods / instruments for all analytes within your Registered Scheme are available to download online.

### Mainline Chemistry Summary Sheet Distribution QH

Distribution: QH							
Distribution Date: 02-Nov-15							
Analyte: Creatinine (µmol/L)							
Method	Instrument	1	2	3	4		
		Overall Mean	352.8	350.0	43.3	583.8	
		Overall SD	13.7	13.9	5.1	25.6	
		Est. Uncertainty of Consensus	0.84	0.85	0.31	1.56	
		Overall Number	268	270	275	269	
		Reference Value ID-GCMS	351.9		44.3	583.2	
Jaffe - IDMS			Method Mean	347.7	345.6	40.9	571.4
			Method SD	13.8	14.0	4.6	24.9
			Est. Uncertainty of Consensus	1.08	1.09	0.36	1.98
			Number	162	163	165	158
		Advia 1200/1650/1800/2400	Instrument Mean	342.4	341.5	37.0	571.8
			Instrument SD	4.7	5.7	3.1	6.5
			Number	16	16	16	15
		AU2700/AU5400/AU5800	Instrument Mean	341.1	338.8	40.8	565.1
			Instrument SD	8.8	8.1	5.1	15.5
			Number	21	20	20	21
		AU400/600/640/680	Instrument Mean	341.9	341.3	38.0	563.5
			Instrument SD	7.1	6.9	1.5	13.6
			Number	16	16	16	16
		Architect	Instrument Mean	367.2	366.2	45.3	617.3
			Instrument SD	5.7	6.5	1.1	8.3
			Number	24	24	24	24
		DX	Instrument Mean	356.3	362.2	37.6	595.8
			Instrument SD	2.4	5.3	1.6	5.2
			Number	6	6	5	6
		Daytona	Instrument Mean	327.3	324.5	39.2	537.0
			Instrument SD	19.3	21.3	7.5	33.7
			Number	13	14	13	14
		Cobas C Module	Instrument Mean	347.6	345.0	41.0	559.8
			Instrument SD	10.1	8.9	2.9	15.3
	Number		59	58	61	57	

## Figure 12 - "End of Batch" report


Material is prepared in a number of Programmes to cover more than one distribution, e.g. a batch of mainline chemistry samples consists of 8 levels and each level is distributed on 4 or 5 occasions over a 10 month period. This allows calculation of your between batch imprecision.

The report provides the mean, SD and coefficient of variation (CV%) of your results for each level over this time period. Your CV is then compared with the median CV of all methods and the median CV of your method.

Lab Code: AE . Section Name: Aeroset 1. Scheme: Mainline Chemistry. Distribution Range: KR - LA

Analyte: Potassium

Method: Indirect ISE	M591	M592	M593	M594	M595	M596	M597	M598	M599	M600	M601	M602
<b>Section Stats</b>												
Mean reported results	1.77	2.23	2.63	3.14	3.63	4.07	4.56	5.11	5.56	6.07	6.59	7.03
SD reported results	0.03	0.01	0.05	0.04	0.03	0.07	0.08	0.08	0.04	0.07	0.02	0.10
CV(%) reported results	1.66	0.52	1.95	1.36	0.89	1.74	1.86	1.48	0.73	1.07	0.23	1.44
Number of results	4	3	3	4	3	2	4	3	3	4	3	3
<b>Method Result Stats</b>												
Mean method mean	1.83	2.30	2.76	3.23	3.71	4.19	4.68	5.16	5.66	6.14	6.63	7.14
Median CV	2.67	2.44	2.01	1.55	1.53	1.35	1.09	1.12	1.02	0.94	0.88	0.83
<b>Overall Result Stats</b>												
Median CV	2.67	2.44	2.00	1.55	1.53	1.36	1.10	1.12	1.02	0.94	0.88	0.84




## Figure 13 - Regional / Network reports

Using the "end of batch" data, reports can also be generated to compare your performance within your region or network. The report provides a tabular and graphical representation of the deviation (absolute or %) from the group mean for each laboratory for each analyte.

The Weqas performance criteria is also provided on the graph for information. These reports are available to print from your browser, or e-mailed in PDF format to your e-mail address.


Absolute Deviation from Regional Means												
	B291	B297	B300	B296	B301	B292	B293	B294	B295	B298	B299	B302
FH Advia 2400 1	-0.59	-9.31	-10.89	-5.39	-14.48	-2.56	-3.67	-5.22	-7.38	-6.15	-10.04	-11.82
FH Advia 2400 2	-1.92	-11.48	-18.52	-9.36	-13.35	-2.92	-4.20	-5.86	-7.84	-5.65	-14.51	-14.05
GF Beta	1.21	5.62	9.58	9.14	8.08	2.24	1.13	4.21	0.06	8.95	0.72	6.72
GF Alpha (4)	1.21	3.49	11.58	7.81	16.42	3.24	2.47	6.88	3.72	11.62	11.72	10.72
KJ ARCHITECT 2	0.88	11.49	8.25	3.48	2.42	1.91 *	1.13	4.21 *	3.39	-3.88	2.06	-0.62
KJ ARCHITECT 1	-0.79	6.49 *	1.25 *	-5.69	0.92	0.91 *	3.13	1.21 *	8.06	-4.88	10.06	9.05
KJ ARCHITECT 1	3.21 *	16.49 *	9.25 *	16.81 *	-	-	-	-	-	-	-	-
Weqas SD	2.31	8.61	12.09	7.39	12.90	3.24	4.27	5.34	6.41	9.47	10.93	14.24

\* - not included in group mean



### Figure 14 - Simplified Reports

An alternative simplified report is available as an alternative to the Standard report. Identical statistical analysis and data evaluation is undertaken, however the report is simplified to colour blocks and a performance alert for non laboratory personnel.




Colour coded graphical representation of Individual SDI scores over 6 month time frame


6 month time frame of bias plot across concentration range.

Performance alert based on Westgard rules

The following rules are used for the Performance alert.



At least 1 result with SDI > 3




At least 2 results with SDI > 2

If number of samples in current distribution ≥ 2 then applies to this dist only.

If number of samples in current = 1 then applies to this and at least one in previous distribution.

**Consistent Bias in one direction.**


This is calculated over several distributions and will depend on the number of samples distributed per round e.g. if 3 samples are distributed per round, then the rule will be calculated over 2 distributions.



# samples in current with analyte	Number of samples overall to consider	trigger
1	6 samples 6 distributions	6 <sub>x</sub> : 6 samples with SDI > 0.2 <i>or</i> 6 samples with SDI < -0.2
2	6 samples 3 distributions	6 <sub>x</sub> : 6 samples with SDI > 0.2 <i>or</i> 6 samples with SDI < -0.2
3	6 samples 2 distributions	6 <sub>x</sub> : 6 samples with SDI > 0.2 <i>or</i> 6 samples with SDI < -0.2
4	8 samples 2 distributions	8 <sub>x</sub> : 8 samples with SDI > 0.2 <i>or</i> 8 samples with SDI < -0.2
5	10 samples 2 distributions	10 <sub>x</sub> : 10 samples with SDI > 0.2 <i>or</i> 10 samples with SDI < -0.2
6	6 samples 1 distribution	6 <sub>x</sub> : 6 samples with SDI > 0.2 <i>or</i> 6 samples with SDI < -0.2

**At least 3 results > 1 SDI.**

This is calculated over several distributions and will depend on the number of samples distributed per round e.g. if only 1 sample is distributed per round, then the rule will be calculated over 3 distributions.



# samples in current with analyte	Number of samples overall to consider	trigger
1	3 samples 3 distributions	3 <sub>1s</sub> : 3 samples with SDI > 1 <i>or</i> 3 samples with SDI < -1
2	4 samples 2 distributions	4 <sub>1s</sub> : 4 samples with SDI > 1 <i>or</i> 4 samples with SDI < -1
3	3 samples 1 distributions	3 <sub>1s</sub> : 3 samples with SDI > 1 <i>or</i> 3 samples with SDI < -1
4	4 samples 1 distributions	4 <sub>1s</sub> : 4 samples with SDI > 1 <i>or</i> 4 samples with SDI < -1
5 or more	All in current 1 distribution	4 <sub>1s</sub> : 4 samples with SDI > 1 <i>or</i> 4 samples with SDI < -1

## 2. Statistical Analysis - Qualitative Programmes

### 2.1 Target Value Assignment

The spiked values are used to determine the target value, verified whenever possible by quantitative analysis. For endogenous samples the result from quantitative analysis is used. When quantitative data is not available, interpretation is based on the majority percentage of responses from participants.

### 2.2 Scoring System

The scores broadly reflect clinical importance. A correct result (in agreement with interpretive comment) is given a score of 0.

A sliding scale score of between 1 and 5 is assigned for incorrectly identified results, where 5 represented a gross misclassification of the result.

A negative result for a positive sample is given a score of 3 to 5 depending on the concentration of the positive sample.

A positive result for a negative sample is given a score of 2 or 3.

Equivocal comments (for further investigation) for a positive sample are given a score of 1 to 3 depending on the concentration of the positive sample.

An equivocal comment (for further investigation) for a negative sample is given a score of 1.

The sensitivities of the methods, the intended purpose of the kits, whether “rule in” or “rule out” are also taken into account in the scoring. In general, a missed positive sample is given a larger penalty than a misclassified negative as this could lead to missed diagnosis or inappropriate treatment whilst an incorrect negative tends to lead to less severe clinical consequences such as inappropriate further investigation.

**Table 9 - Qualitative Scores**

Lab Result	Target Value	Score
+ve	+ve	0
equivocal	+ve	1, 2 or 3
-ve	+ve	3, 4 or 5
-ve	-ve	0
equivocal	-ve	1
+ve	-ve	2 or 3

Individual sample scores are added together and averaged for the distribution to provide an overall analyte score. However, a negative for a negative result score of 0 is not included in the overall analyte score.

**Table 10 - Interpretation of Scoring System**

When the individual score is:

Score	Interpretation
0	Good
1	Acceptable
2	Warning
> 2	Unacceptable

These Scores are treated in the same way as SDI scores for Performance surveillance. Please refer to Section 6, Performance Surveillance.


## 2.3 The Weqas Report

An example of a typical participant's report for the Pregnancy Testing scheme is given below. Each report includes the scoring criteria, a summary of the qualitative results, the broad method used (manufacturer), and method specific performance.

Figure 15 - Manager's Summary Report

Lab: AE . Scheme: Urine Pregnancy Testing, Distribution Code: W68.  
Final Report Issued: 5/10/11

office@weqas.com  
Scheme Organiser:  
Annette Thomas



### Section SDI scores for this distribution

Section	2TB2	Clinical Research Facility	Dermatology	EAU - Lisa Waters	EAU - Medical A1 Link	EAU - Surgery	Emergency Gynae	FP Broad Street	FP Butetown
<b>Overall</b>	<b>1.00</b>	<b>0.00</b>	<b>0.00</b>				<b>0.00</b>	<b>0.00</b>	
Qualitative HCG (High Sensitivity)	1.00 (avg)	0.00 (avg)	0.00 (avg)	?	?	?	0.00 (avg)	0.00 (avg)	?
Section	FP Cardiff Royal	FP Gabalfa	FP Grangetown	FP Heath, C/O ANC	FP Llanrunney	FP Llantwit	FP Park View	FP Penarth	FP Roath
<b>Overall</b>	<b>0.00</b>	<b>1.00</b>	<b>0.00</b>	<b>0.00</b>	<b>0.00</b>	<b>0.00</b>		<b>0.00</b>	<b>0.00</b>
Qualitative HCG (High Sensitivity)	0.00 (avg)	1.00 (avg)	0.00 (avg)	0.00 (avg)	0.00 (avg)	0.00 (avg)	?	0.00 (avg)	0.00 (avg)

SDI Code	Meaning
N/A	Not enrolled for this analyte
?	Analyte enrolled but no results returned
N/S	This analyte not scored
**	SDI score greater than 2

Please note: Method and Instrument Summary reports are available to download via the 'Lab Stats' or 'Section Stats' menu.  
If you don't currently have interactive access, please contact WEQAS for a registration form on 02920 314750.

Comments:

	Sample 1	Sample 2	Sample 3
<b>Urine Source</b>	Urine from non pregnant donor	Urine from pregnant donor diluted to approx 29iu	Urine from pregnant donor diluted to approx 336iu
<b>Interpretation</b>	Negative	Weak Positive	Positive

For interpretation purposes, a sample is regarded negative at a concentration less than 20 IU/L (equivocal results may be produced at a concentration range of 10-20 IU/L and therefore no penalty is given for returning a positive or weak positive result in this equivocal range.) However reporting positive results for a concentration of < 10 IU/L will incur a penalty.  
A sample is regarded positive at a concentration >20 IU/L.

## Figure 16 - Individual Selection Report

The individual section report includes a graphical representation of the participant's results compared with other participants using the same method (white bar), results for all methods (grey bar) and the correct interpretation based on the quantitative result (green bar). In the absence of a quantitative result the correct interpretation is based on the majority percentage of responses from participants.

### Qualitative Report

Lab Code: AE Section: 2TB2

### Qualitative HCG (High Sensitivity) Results

Lab Code	Section	Method	Instrument	Sample Number			Sample Score			Average Score (Average)
				1	2	3	1	2	3	
AE	2TB2	Unipath	Clearview HCG (3min)	Negative	Negative	Positive	0	2	0	1.00
<b>Interpretation</b>				<b>Negative</b>	<b>Wk Positive</b>	<b>Positive</b>				
<b>Spiked Value</b>				<b>Urine from non pregnant donor</b>	<b>Pregnant donor urine diluted to approx 29iu</b>	<b>Pregnant donor urine diluted to approx 336iu</b>				




Figure 17 - Example of kit summary page - available with every distribution.

Qualitative Report

Distribution W68

Qualitative HCG (High Sensitivity) Results



Method="Unipath" [Show all methods](#)

Lab Code	Section	Instrument	Sample Number			Sample Score			Average Score
			1	2	3	1	2	3	
IJ	Clinical Pathology	<a href="#">Clearview HCG (5min)</a>	Negative	Equivocal,Borderline	Positive	0	1	0	0.50
ABE	Clearview EasyHCG	<a href="#">Clearview Easy HCG</a>	Negative	Wk Positive	strong positive	0	0	0	0.00
ADL	EASY HCG IV502575	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	A&E UCH	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	AAU UCH	<a href="#">Clearview Easy HCG</a>	Negative	Negative	Negative	0	2	3	2.50
JO	Archway Sexual Health	<a href="#">Clearview Easy HCG</a>							
JO	Basil Samuel OP NHNN	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	Cardiac Catheter	<a href="#">Clearview Easy HCG</a>							
JO	Clin Biochem	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	Clinic K UCH	<a href="#">Clearview Easy HCG</a>							
JO	Clinical Research UCH	<a href="#">Clearview Easy HCG</a>	Positive	Negative	Positive	2	2	0	1.33
JO	Day Surgery	<a href="#">Clearview Easy HCG</a>	Negative	Negative	Positive	0	2	0	1.00
JO	Diagnostic Gynae Unit	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	EGA Breast Clinic	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	EGA Repro Med Unit	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	Hyperacute Stroke Unit UCH	<a href="#">Clearview Easy HCG</a>							
JO	Lady Ann Alerton NHNN	<a href="#">Clearview Easy HCG</a>							
JO	Mortimer Market Centre	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	MRI unit, NHNN	<a href="#">Clearview Easy HCG</a>	Negative	Wk Positive	Positive	0	0	0	0.00
JO	National Day Care RLHH	<a href="#">Clearview Easy HCG</a>	Negative	Negative	Positive	0	2	0	1.00
JO	Nuclear Medicine	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	Nuffield Ward NHNN	<a href="#">Clearview Easy HCG</a>							
JO	OncoChemo	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	OP HTD	<a href="#">Clearview Easy HCG</a>	Negative	Wk Positive	Positive	0	0	0	0.00
JO	OPD Derm 5th Fl	<a href="#">Clearview Easy HCG</a>							
JO	POPD EGA	<a href="#">Clearview Easy HCG</a>	Negative	Positive	Positive	0	0	0	0.00
JO	T10	<a href="#">Clearview Easy HCG</a>							
GU	Rapid Assessment Unit	<a href="#">Clearview HCG (3min)</a>							
GU	Samaritan OPD	<a href="#">Clearview HCG (3min)</a>							
GU	Samaritan Ward	<a href="#">Clearview HCG (3min)</a>							
GU	Surgery	<a href="#">Clearview HCG (3min)</a>							
GU	Urogynaecology	<a href="#">Clearview HCG (3min)</a>							
GU	Winstan 1st Floor	<a href="#">Clearview HCG (3min)</a>							
IM	A+E Garrick	<a href="#">Clearview HCG (3min)</a>	Negative	Negative	Positive	0	2	0	1.00
IM	HCG/main lab	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
IM	POCT 1	<a href="#">Clearview HCG (3min)</a>							
IM	POCT 2	<a href="#">Clearview HCG (3min)</a>							
IM	POCT 3	<a href="#">Clearview HCG (3min)</a>							
IM	POCT 4	<a href="#">Clearview HCG (3min)</a>							
IM	POCT 5	<a href="#">Clearview HCG (3min)</a>							
IM	Satellite	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
IV	Haematology	<a href="#">Clearview HCG (3min)</a>	Negative	Equivocal,Borderline	Positive	0	1	0	0.50
ML	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Equivocal,Borderline	Positive	0	1	0	0.50
MO	HCG	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
MP	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Negative	Positive	0	2	0	1.00
MS	Eland Ward (First Floor)	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
MS	Leeds Daycare	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
MS	Leeds Ward 2	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
MS	Longland (Methley Ward)	<a href="#">Clearview HCG (3min)</a>	Negative	Negative	Positive	0	2	0	1.00
MS	Methley Ward	<a href="#">Clearview HCG (3min)</a>	Negative	Wk Positive	Positive	0	0	0	0.00
MJ	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
NN	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
OM	microbiology / pregnancy	<a href="#">Clearview HCG (3min)</a>							
SP	Casualty (Clearview)	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
SP	Laboratory (Clearview)	<a href="#">Clearview HCG (3min)</a>	Negative	Positive	Positive	0	0	0	0.00
VK	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Equivocal,Borderline	Positive	0	1	0	0.50
YH	Pregnancy	<a href="#">Clearview HCG (3min)</a>	Negative	Wk Positive	Positive	0	0	0	0.00
YN	MICROBIOLOGY PREG	<a href="#">Clearview HCG (3min)</a>	Negative	Wk Positive	Positive	0	0	0	0.00
ABE	TestPack 505798	<a href="#">Test Pack</a>	Negative	Wk Positive	strong positive	0	0	0	0.00
ABE	TestPack 505805	<a href="#">Test Pack</a>	Negative	Wk Positive	strong positive	0	0	0	0.00

Interpretation	Negative	Wk Positive	Positive
Spiked Value	Urine from non pregnant donor	Pregnant donor urine diluted to approx 29iu	Pregnant donor urine diluted to approx 336iu



## INTERPRETATION OF LABORATORY EQA REPORTS

# Weqas



+44 (0)2920 314750



+44 (0)2920 314760



contact@weqas.com



www.weqas.com



Weqas  
Unit 6, Parc Tŷ Glas  
Llanishen  
Cardiff, UK  
CF14 5DU



4301



4301