

Version: 1.1 PL-QL1-RptIntStandard Page 1 of 3 © Copyright Weqas

'Your reported result' will be shown in the units entered, in the Weqas standard units if units entered are different from the Weqas standard units, and method corrected (if appropriate for the

analyte).

Target Value: This will show whether you have been scored against Reference Value, Method Mean or Overall Mean.

Wegas TAE: This is the Total Allowable Error (2 x Wegas SD).

PI (Performance Index) score is shown for each sample.

 $PI = \frac{\text{Reported Result - Target Value}}{\text{(Wegas TAE / 2)}}$

Overall performance category is assigned using an algorithm based on the individual sample PIs.

The Bias (Absolute) Plot shows absolute bias in Weqas standard units for each sample for the current distribution. The x-axis is the target value; if the target value is a Reference Value or the Overall Mean, your Method Means will be plotted.

Programme: ED Toxicology • Distribution Code: SP0622

Distribution Start: 30-May-2022 • Distribution End: 15-Jun-2022 • Report Issued: • Report Status: n/a


Participant Code: WQ00000 • Location: Main Lab • Analyser Name: n/a • Serial #:


Date Samples received: 30-May-2022 • Date of Analysis: 09-Jun-2022 • Operator Details: ## • Storage Conditions: 4 °C


Analyte: Salicylate • Method: Roche Enzymatic • Kit: SALI

		Salio	ylate mg/L		
	Your reported result (mg/L)		Method: Roche Enzymatic	Instrument Model: cobas c 501	Overall
		Mean	265.4	270.1	268.5
Sample 1	264.3	SD	5.6	5.4	9.3
ounipie 1	200	Uncertainty	1.23	2.27	1.58
		n	33	9	54
	605.8	Mean	610.9	618.6	622.8
Sample 2		SD	12.2	12.8	23.6
		Uncertainty	2.66	5.32	4.01
		n	33	9	54
		Mean	352.4	356.7	356.6
Sample 3	349	SD	6.4	6.7	11.7
		Uncertainty	1.39	2.81	2.00
		n	33	9	54
Sample 4	90.7	Mean	92.7	91.8	95.1
		SD	2.2	1.9	4.8
		Uncertainty	0.48	0.77	0.82
		n	33	9	54

Scoring					
Sample	1	2	3	4	Overall Performance
Target: (Method Mean)	265.4	610.9	352.4	92.7	
Weqas TAE	19.1	54.7	25.8	10.4	
PI	-0.12	-0.19	-0.26	-0.39	Good







Precision	
	This distribution: SP
Sy.x	1.12
S	0

Sy.x is the average deviation from the best fit line and is a
index of scatter.

Interpretation
Good
Acceptable to Warning level
Unacceptable (incl. Curvilinear Data)


Running PI chart shows the PI score for each sample for the current distribution and the last 5 distributions, with colour coding for each performance category.


The Bias (Relative) Plot show the relative (%) bias to the target for the current results, and for the previous 5 distributions.


Version: 1.1 PL-QL1-RptIntStandard Page 2 of 3 © Copyright Weqas

The histogram shows breakdown of results for each sample, showing the spread of results for All Results, 'My Method' and 'My Instrument'.

The Method Summary shows Mean and CV for each Method for each sample.







The orange dot shows where your result lies.

■ All Results ■ My Method ■ My Instrument • My Result

Method Summary

SP0622	Samp	Sample 1		Sample 2		Sample 3		Sample 4	
Method Name	Mean	CV%	Mean	CV%	Mean	CV%	Mean	cv%	
Abbott Enzymatic (n=4)	275.8	0.6	634.0	3.0	365.3	2.3	98.5	3.8	
Beckman EMIT (n=9)	271.9	7.1	659.9	10.1	371.3	10.1	98.8	6.5	
CLS Enzymatic (n=1)	264.6	0.0	635.8	0.0	357.9	0.0	93.8	0.0	
Roche Enzymatic (n=33)	265.4	2.1	610.9	2.0	352.4	1.8	92.7	2.4	
Siemens EMIT (n=2)	262.0	0.6	646.5	1.1	355.5	6.1	95.0	10.1	
Siemens Enzymatic (n=5)	295.4	3.4	686.0	4.2	396.8	4.4	110.6	4.9	

There are no Weqas or Participant supplied comments for Salicylate for this distribution.

Any analyte specific comments from Wegas or the participant are shown.

Version: 1.1 PL-QL1-RptIntStandard Page 3 of 3 © Copyright Weqas